<u>NATIONAL INSTITUTE OF FASHION TECHNOLOGY</u> (NDII Project – Head Office)

No. 1351(37)NIFT/Pur-HO/Tensile Strength/2020

2nd November, 2020

Addendum

This is with reference to tender document no. 1351(37)NIFT/Pur-HO/Tensile Strength/2020 dated 16th October2020; the following addendum may be noted please

1. Temperature conditions in specifications (1b, civ & d2) may be read as per International and Indian Standards

i.e 20 to 27 degree C±2 $^{\rm o}$ C.

2. Rest will be same as per tender document.

3. If any bidder has uploaded the bid against the said tender as per previousspecifications before uploading the revised specifications, they can upload the revised price bid on e-portal accordingly.

4. This issues with the approval of Competent Authority of NIFT.

Assistant Director (Purchase) NIFT – Head Office